

MMO: MEA

CALIBRATION REPORT

Document under Configuration Control




Prepared by Andrey Fedorov	Date	Signature:
Verified by Claude Austin		Signature:
Approved and Application authorized by: Jean Andre Sauvaud		Signature:

Summary : MEA FM calibration report and description of the calibration data record

Keywords : Calibration

Annex: Calibration Data




I DISTRIBUTION LIST

PROJECT	EXTERNAL DISTRIBUTION
C. Aoustin E. Le Comte J.-A. Sauvaud J. Rouzaud P. Rouger Q.-M. Lee H.-C. Seran M. Petiot C. Garat D. Moirin	Y. Saito

II LIST OF ABBREVIATIONS

MMO	Mercury Magnetosphere Orbiter		
MEA	Mercury Electron Analyzer		
TBC	To Be Completed		
TBD	To Be Defined		
TBW	To Be Worked		



III DOCUMENT CHANGE RECORD

Edition	Revision	Date	Modified pages	Reason for change / Observations
0	0	04/12/2012		First drafty version
1	0	25/01/2013	all	Preliminary version
2	0	15/03/2013	all	Complete version





TABLE OF CONTENTS

MMO: MEA	1
1. Scope of this document.....	5
2. Definitions.....	5
3. MEA theoretical properties.....	6
4. Calibration setup	7
4.1 Mechanical setups	7
4.1.1 Full calibration setup.....	7
4.1.2 MCP characterization setup	8
4.1.3 UV contamination test setup	8
4.2 Beam properties	9
5. Calibration results and data record	9
5.1 MCP characterization	9
5.1.1 MCP image.....	9
5.1.2 MCP count profile.....	10
5.1.3 MCP noise	10
5.1.4 MCP efficiency variations.....	11
5.1.5 MCP dead time.....	11
5.2 MEA full calibration	11
5.2.1 Full calibration procedure.....	11
5.2.2 Full calibration results	12
5.3 UV contamination results.....	15
6. Summary of the calibration results.....	16
ANNEX 1 SUPPLEMENTARY DATA FILES	17
ANNEX 2 ELECTRON BEAM CALIBRATION.....	18



1. Scope of this document

This document contains a complete description of the MEA calibration setup, MEA calibration procedure and MEA calibration data. The document shows also the theoretical properties of the conceptual design to make a comparison the conception properties and the real instrument properties. To understand this document, reader should refer to the “MEA Instrument Design Report”. The Annexes contains the tables of the supplementary data files and the electron beam calibration details.

2. Definitions

- E Incident electron energy, eV
- θ Elevation angle (see Figure 1,2)
- Φ Azimuth angle (see Figure 1,2)
- U_{AN} Analyzer voltage
- U_{TOP} Top part of analyzer voltage
- K $E = U_{AN} \cdot K$
- $K0$ The best K for the current θ and Φ
- $\Delta E/E$ Energy resolution of the analyzer
- P_{BEAM} electron beam flux $cm^{-2} s^{-1}$ as a function of the elevation angle
- Ω_i one azimuthal sector aperture, cm^2 for fixed θ , Φ , U_{AN} , and U_{TOP}
- C_i count rate, s^{-1} of one azimuthal sector
- G_i one sector geometrical factor, $cm^2 sr eV/eV$
- G the total geometrical factor of the instrument, $cm^2 sr eV/eV$ (used for numerical simulation)
- HV_{MCP} The MCP high voltage, V, measured at the HV unit level.

Figures 1,2,3 show the MEA axis definition, spherical frame of measurement definitions, azimuthal sectors (anodes) numbering and their position in the MEA frame and in the MMO frame. Note that Figure 2 gives looking directions of the anodes.




Figure 1 MEA frame and spherical angles





Figure 2 MEA azimuth sectors (anodes) looking directions in the MMO frame



3. MEA theoretical properties

MEA theoretical properties were obtained with a realistic 3-D numerical model.

The results of ray tracing are shown in Figures 3,4. The summary of the numerical results is in Table 1 in the bottom of this section. Technically, the instrument response was obtained as follows:

1. Set constant U_{AN} , U_{TOP}
2. Illuminate the sensor by the parallel uniform electron beam with variable:
 - (a) Energy
 - (b) Elevation angle
 - (c) Azimuth angle
3. Calculate Ω_i of the sensor for each value of the energy, azimuth, and elevation
4. Do the same for several U_{TOP} / U_{AN} values, that should represent the set of geometrical factors of the instrument.

Figure 3 represents the variation of the instrument responses with geometrical factor (note that elevation angle definition here is opposite to the Calibration Elevation angle, Figures 1 and 13). Figure 4 shows the profile of G as a function of U_{TOP} / U_{AN} .




Figure 3 Elevation (left) end energy responses of MEA for different geometrical factors




Figure 4 G as a function of U_{top}/U_{an}

The corresponding best fit formula is:

$$G [cm^2 sr eV/eV]$$

$$= -15.05 + 32.65 \cdot \left(\frac{U_{Top}}{U_{An}}\right) - 36.18 \cdot \left(\frac{U_{Top}}{U_{An}}\right)^2 + 19.14 \cdot \left(\frac{U_{Top}}{U_{An}}\right)^3 - 5.31 \cdot \left(\frac{U_{Top}}{U_{An}}\right)^4$$

#	U_{TOP} / U_{AN}	K	$\Delta E / E$	$G, cm^2 sr eV/eV$
1	1.1	10.2	0.09	$0.51 \cdot 10^{-2}$
2	1.0	9.95	0.10	$0.83 \cdot 10^{-2}$

3	0.9	9.75	0.11	$1.03 \cdot 10^{-2}$
4	0.8	9.56	0.11	$1.15 \cdot 10^{-2}$
5	0.65	9.37	0.13	$0.78 \cdot 10^{-2}$
6	0.57	9.34	0.15	$0.51 \cdot 10^{-2}$
7	0.52	9.34	0.14	$0.37 \cdot 10^{-2}$
8	0.4	9.25	0.10	$0.12 \cdot 10^{-2}$
9	0.3	9.03	0.07	$0.03 \cdot 10^{-2}$

Table 1 Summary of MEA numerical simulation

4. Calibration setup

4.1 Mechanical setups

The IRAP big vacuum chamber was used for MEA calibration. There are three type of tests and correspondingly three types of mechanical setups:

- Full calibration setup
- MCP characterization setup
- UV contamination test setup

4.1.1 Full calibration setup

The mechanical MEA setup is shown in Figure 5. The center of the beam is matched to the line of sight of the instrument. The θ rotation center is very close to its line of sight. The corresponding photo is shown in Figure 6. This figure shows as well the beam monitor and the magnetometer dedicated to the run time reading of the magnetic field in the vicinity of the instrument.




Figure 5 MEA full test mechanical setup. Yellow rectangular shows the electron beam size




Figure 6 MEA-1 in the vacuum chamber. A magnetometer is glued to the sensor housing. The instrument is turned on 90° in elevation. Note the thermal shield simulator above the instrument aperture.

To make the calibration results relevant, the calibration has been made with a realistic simulator of the instrument thermal shield. The Earth's magnetic field inside the vacuum chamber is compensated by the Helmholtz coils. Continuous measurements of the residual magnetic field and its run-time automatic adjustment allows to keep the maximal value of the magnetic field about $0.5\mu T$. See the beam properties in Section 4.2 and Annex 1.

4.1.2 MCP characterization setup

MCP characterization setup is quite different for the full calibration one:

- a. The sensor is disassembled to expose MCP directly to the beam (see Figure 7);
- b. The special attractive diaphragm is installed around the MCP ring to avoid contamination of the measurements by secondary electrons (not shown). The diaphragm potential is +5V.
- c. The MCP is tested by a needle-like electron beam generated by a special Y-Z mobile electron gun (Figure 7)




Figure 7 Left: The sensor with no analyzer, Right: the MCP characterization setup. One can see the needle-like beam electron gun.

4.1.3 UV contamination test setup

UV contamination test has been made with HAMAMATSU Deuterium lamp L7292. The setup is shown in Figure 8. The beam photon rate in the VUV range ($\sim 160\text{nm}$) at the instrument level is $2.4 \cdot 10^{11} \text{ s}^{-1} \text{ cm}^{-2}$.





Figure 8 UV test setup

4.2 Beam properties

For each calibration stage the electron beam absolute measurements are performed. The characteristic beam properties are shown in Figure 9. To increase the beam intensity, the UV lamp was located close to the photocathode. This leads to a non-uniformity of the beam as it is shown in Figure 9. Such a distribution is measured by scanning the beam by a dedicated monitor unit with a pin-hole entrance (Figure 6). At the same time we measure the total electron current leaving the photocathode. The monitor count rate is converted to the real flux assuming that the total count of a 2D distribution corresponds to the total electron current of the photocathode. The beam flow P_{BEAM} for each elevation position of the sensor was recalculated as it is shown in Figure 9 bottom. This distribution is calculated by averaging of the original 2D distribution inside a rectangle $12 \times 5mm$. The position of this rectangle in the 2D beam flux distribution is calculated from the mechanical setup shown in Figure 5. The beam distribution was measured for each calibration run. The beam measurements corresponding to each calibration run are shown in Annex 2.




Figure 9 Electron beam properties

5. Calibration results and data record

5.1 MCP characterization

5.1.1 MCP image

MCP image have been measured to provide information about MCP heals and MCP efficiency uniformity. To do this the MCP has been scanned by a needle-like electron beam. The results of such scan for MEA-1 and MEA-2 are shown in Figure 10. The distributions show that all anodes (except MEA-1 closed anodes are functional, and there is not significant efficiency difference.





Figure 10 Left: MEA-1 MCP image, Right: the same for MEA-2

5.1.2 MCP count profile

MCP anodes counts as a function of HV_{MCP} are shown in Figure 11. The measurements were done with a needle-like beam positioned in the anodes gravity centres (points marked by stars in Figure 10, left). Note that HV_{MCP} gives the voltage at the HV unit output. The real MCP bias is 400V less. The working point of both MCPs (MEA-1 and MEA-2) is chosen as 2750V. The corresponding text data of the calibration package see in Annex 2.




Figure 11 MCP normalized count as a function of the HV_{MCP}

5.1.3 MCP noise

MCP noise has been measured for MEA-1 only since there are completely obscured anodes. The results are shown in Figure 12. See the corresponding text file in Annex 1.





Figure 12 One anode MCP noise versus HV_{MCP}



5.1.4 MCP efficiency variations

MCP anode-by-anode efficiency variation was made at the same time as the count v. HV_{MCP} . The measurement conditions are as follows:

- Electron beam: 1000eV
- $HV_{MCP} = 2750 V$

The results are shown in Figure 13. Corresponding best fit formulas are as follows:

$$\begin{aligned} Rel. Count &= 0.10 \cdot \sin((N_{Anode} - 4.02) \cdot 4\pi/16) & \text{MEA-1} \\ Rel. Count &= 0.22 \cdot \sin((N_{Anode} - 4.64) \cdot 4\pi/16) & \text{MEA-2} \end{aligned}$$




Figure 13 Relative anodes efficiency

The corresponding text files are described in Annex 1.

5.1.5 MCP dead time

The MCP dead time of one anode was measured by defocused beam with a known ratio between the intensity in the beam center and beam periphery. Thus, varying, the beam flux, we have a linear count at one anode and saturated count in another anode. The results are shown in Figure 14. The one anode saturation level is the same for the both sensors. The MCP-amplifier chain saturation can be describes by formula:

$$C_i = CO_i / (1 + CO_i \cdot \tau)$$

Here CO_i is an incident electron rate of one anode and τ is defined from the table:

	MEA-1	MEA-2
$\tau, \mu s$	1.2	1.22

The corresponding data file see in Annex 1.




Figure 14 MCP saturation profiles

5.2 MEA full calibration

5.2.1 Full calibration procedure

The full MEA-1 and MEA-2 calibration was made in the coordinate frame shown in Figure 15. The beam properties is as follows:

Item	Value	Comment
Beam Energy	1000 eV	To avoid any influence of the residual magnetic field
Beam number flux density MEA 1	about $1.7 \cdot 10^5 \text{ cm}^{-2} \text{ s}^{-1}$	See Annex 2
Beam number flux density MEA 2	about $2.4 \cdot 10^5 \text{ cm}^{-2} \text{ s}^{-1}$	See Annex 2

Table 2 Full calibration run conditions




Figure 15 Calibration angles definition

A brief description of the calibration scheme:

Dimension	Range	Step	N steps
Azimuth, Φ	$-166^\circ \div +166^\circ$	2°	167
Elevation, Θ	$-8^\circ \div +18^\circ$	1.5°	18
Analyzer voltage, $\Delta K/K_0$	$0.78 \div 1.24$	0.025	15
Top part voltage MEA-1, U_{TOP} / U_{AN}	{1.0, 0.8, 0.52, 0.37, 0.28}		5
Top part voltage MEA-2, U_{TOP} / U_{AN}	{1.0, 0.8, 0.42, 0.34, 0.27}		5

Table 3 The full calibration scheme

For each Azimuth angle Φ , we make a scan over elevation angle θ . For each angular position we test the set of geometrical factors values. For each value of geometrical factor attenuation we scan the analyzer voltage.

5.2.2 Full calibration results

The main raw result of the full calibration could be expressed as an anode effective aperture (Ω_i) as a function of azimuthal angle Φ , elevation angle θ , and energy (U_{AN} actually) for all geometrical factor attenuation levels. The raw data converted from counts to Ω_i are presented in a self-described text file. See Annex 1 for details. The integral of these data over the energy and θ are shown in Figure 16. One can see the modulation of the geometrical factor as a function of beam azimuth increasing when geometrical factor is reducing. The modulation of K factor as a function of the anode number (azimuth) is shown in Figure 17. The variation of the central angle θ as a function of the anode number (azimuth) and the G level is shown in Figure 18. Figures 19 and 20 show the energy and elevation responses for different anodes and different G levels. The dashed curves in these plots shows the polynomial fit that allows to define the $\Delta E/E$ and $\Delta \theta$ with high accuracy. Anodes geometrical factor versus the U_{TOP}/U_{an} value is shown in Figure 21. The theoretical profile obtained from the numerical simulation is shown by dashed curve. The attenuation factor values, that include the grid transparencies and MCP efficiency are as follows:

	MEA-1	MEA-2
Theory/calibration factor	20 x 4.0	5.5

The corresponding final values, as G_i , K , θ_0 , $\Delta E/E$, $\Delta\theta$, etc, applicable for the data analysis are given in the dedicated text files. See Annex 1 for details.







Figure 16 Azumuthal density of one anode geometrical factor




PS file name: K_AZ_Mea1FMFullTest_1000eV_03Jul_2012.eps
Processed by ShowMeaProp1D.pro.10.0 : Mon Mar 4 11:48:59 2013
Measured Date: 03.07.2012
MCP bias = 2778. V
Utop/Uan=1.00
Utop/Uan=0.80
Utop/Uan=0.52
Utop/Uan=0.37
Utop/Uan=0.28




PS file name: K_AZ_Mea2FMFullTest_19Apr_2012.eps
Processed by ShowMeaProp1D.pro.10.0 : Thu Mar 14 06:42:35 2013
Measured Date: 19.04.2012
MCP bias = 2792. V
Utop/Uan=1.00
Utop/Uan=0.80
Utop/Uan=0.42
Utop/Uan=0.34
Utop/Uan=0.27

Figure 17 Factor K variation along G levels as a function of azimuth and anode number



PS file name: El_AZ_Mea1FMFullTest_1000eV_03Jul_2012.eps
Processed by ShowMeaProp1D.pro.10.0 : Mon Mar 4 11:59:40 2013
Measured Date: 03.07.2012
MCP bias = 2778. V
Utop/Uan=1.00
Utop/Uan=0.80
Utop/Uan=0.52
Utop/Uan=0.37
Utop/Uan=0.28

MEA-1



PS file name: El_AZ_Mea2FMFullTest_19Apr_2012.eps
Processed by ShowMeaProp1D.pro.10.0 : Thu Mar 14 07:08:31 2013
Measured Date: 19.04.2012
MCP bias = 2792. V
Utop/Uan=1.00
Utop/Uan=0.80
Utop/Uan=0.42
Utop/Uan=0.34
Utop/Uan=0.27

MEA-2

Figure 18 Central elevation angle as a function of the G level and azimuth





Figure 19 Energy resolution of different anodes for various G levels (color)




Figure 20 Elevation response of different anodes for various G levels (color)




Figure 21 Anodes geometrical factors as a functions of U_{TOP}/U_{an} . The dashed lines show the theoretical profile scaled with appropriate factor (see the text).

5.3 UV contamination results

UV contamination test results are shown in Figure 22. The strong count at the small energies show the photoelectrons emitted inside the instrument an in the vacuum chamber. One can see that the maximal background is less than 1 per second per anode.




Figure 22 UV contamination tests results

6. Summary of the calibration results

The anodes average of the calibration results are shown in Tables 4 and 5. These data are not applicable for the data processing and should be used for reference purposes.

U_{TOP}/U_{Analyz}	1.0	0.8	0.42	0.34	0.27
$GF \text{ cm}^2 \text{ sr eV/eV}$	$1.01 \cdot 10^{-4}$	$1.71 \cdot 10^{-4}$	$1.61 \cdot 10^{-5}$	$6.32 \cdot 10^{-6}$	$2.17 \cdot 10^{-6}$
GF_0/GF		1.0 (1)	10.65 (6)	27.09 (20)	78.80 (60)
$\Theta \text{ deg}$	2.3	4.8	10.5	11.6	12.5
$\Delta \Theta \text{ deg}$	5.35	5.77	3.45	2.98	2.73
K	9.17	8.73	8.64	8.65	8.61
$\Delta E/E$	0.095	0.119	0.121	0.087	0.044

Table 4 MEA-2 calibration summary


U_{TOP}/U_{Analyz}	1.0	0.8	0.52	0.37	0.28
$GF \text{ cm}^2 \text{ sr eV/eV}$	$5.75 \cdot 10^{-6}$	$9.80 \cdot 10^{-6}$	$3.10 \cdot 10^{-6}$	$8.27 \cdot 10^{-7}$	$3.29 \cdot 10^{-7}$
GF_0/GF		17.46 (20)	55.21 (60)	207 (250)	520 (1000!)
$\Theta \text{ deg}$	0.3	3.0	7.4	9.6	10.8
$\Delta \Theta \text{ deg}$	5.39	5.89	3.86	3.32	2.73
K	8.75	8.35	8.16	8.30	8.29
$\Delta E/E$	0.123	0.155	0.158	0.115	0.074

Table 5 MAE-1 Calibration summary

ANNEX 1 SUPPLEMENTARY DATA FILES

The calibration data package is in the file [MEA_FM_Calibration_Data_2_0_15_Mar_2013.zip](#) attached to this document. The calibration data package consists of the following files:

What is it?	MEA-1 file	MEA-2 file
MCP count as a function of HV_{MCP}	MEA1_FM_MCP_Count_Voltage.txt	MEA2_FM_MCP_Count_Voltage.txt
MCP one anode background count rate as a function of HV_{MCP}	MEA1_FM_MCP_Noise.txt	No data
MCP relative efficiency	MEA2_FM_MCP_Rel_Eff.txt	MEA2_FM_MCP_Rel_Eff.txt
MCP saturation profile	MEA1_FM_MCP_Satur.txt	MEA2_FM_MCP_Satur.txt
Full calibration original data	Mea1FMFullTest_1000eV_03Jul_2012.txt	Mea2FMFullTest_1000eV_19Apr_2012.txt
Full calibration applicable data	Mea1_FM_Properties_1000eV_03Jul_2012.txt	Mea2_FM_Properties_1000eV_19Apr_2012.txt



ANNEX 2 ELECTRON BEAM CALIBRATION

The beam profile as a function of MEA θ taken on 18 Apr 2012 for the MEA-2 full calibrations is:




Figure 23 Beam profile on 18 Apr 2012

The beam profile as a function of MEA θ taken on 11 Jun 2012 for the MEA-1 full calibrations is:




Figure 24 Beam profile on 11 Jun 2012